

Conductivity meters

Cond 7 Vio

Code: 50110422



TECHNICAL DATA SHEET

Standards applied

Regulations

Performance

Accuratezza COND	±2 % f.s.
Accuracy (TDS)	±1 % f.s.
Temperature accuracy	±0,5 °C °C
Temperature coefficient (max)	0,00 ... 10,00 % / °C % / °C
Temperature compensation	0,0 ... 100,0 % / °C °C
TDS factor	0.40 ... 1.00
Measuring range COND	0.00 µS ... 200 mS
TDS measuring range	0.1 mg/l ... 200 g/l
Temperature measuring range	0.0 ... 100.0 °C °C
COND Resolution	0,01 / 0,1 / 1 (scala automatica) µS/cm µS/cm - mS/cm
Temperature Resolution	0.1 °C °C
Working temperature	0 ... +45 °C °C
Reference temperature	+15 ... +30 °C °C

Features

Display	0,1 - 1 - 10
Display	High-definition, backlit colour LCD
IP rating	57 IP
Inputs	BNC - RCA
Calibration points Cond	1 ... 4 automatic values - 1 user value
Calibration Report	Yes
Swabs recognised Cond	84 µS - 147 µS - 1413 µS - 12,88 mS - 111,8 mS
Cell type	2 rings

Feeds

Via della Meccanica, 25, 41012 Carpi (MO) - Italy

+39 059 653274 - info@giorgiobormac.com

Power supply

3 x AA batteries 1.5 V V

Dimensions and weight

Instrument weight

0.4 Kg kg
